

JUN 03 2009

VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130295</u> ✓	
Name <u>JOHN BENNETT</u>		ID# <u>4084</u>	Date <u>6-2-09</u> ✓
Agency <u>APD</u>		Phone # <u>786-8900</u>	
Instrument Location <u>SEELY SUBSTATION - SPENARD RD.</u> ✓			
Alco S/N <u>X301563</u> ✓		Target Value <u>.079</u> ✓	
		High Pressure <u>1200</u>	
Alco Test Values			
<u>.082</u> ✓		<u>.082</u> ✓	
1 st Alco		2 nd Alco	
Signature <u>John Bennett</u>		# <u>27682</u>	
		CORB 6/5/09	

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 12 day of June, 2009.

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130295 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130295

JUNE 02, 2009 ✓

OPERATOR'S NAME:

BENNETT, JOHN

OPERATOR'S NUMBER: 4084

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 1234567

DEPT/AGENCY: ANCI

CASE/REPORT: VOC

TEST TYPE: V ✓

ALCO TARGET VALUE: .079 ✓

ALCO S/N: X301563 ✓

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.40 in ✓
ALCO TARGET .080 14:50
BLANK TEST .000 14:51
INTERNAL STANDARD VERIFIED 14:51
ALCO TV 30.42 in .082 ✓ 14:51
BLANK TEST .000 14:52
SUBJECT SAMPLE .000 14:53
BLANK TEST .000 14:53
ALCO TV 30.40 in .082 ✓ 14:54
BLANK TEST .000 14:54

JP #27682

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130295

JUNE 02, 2009 ✓

TIME 14:57

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 41c
BAROMETER: 30.42 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ Δ

JP #27682